

RETURN AND REPAIR FORM



Customer Information

B:

S:

Company Name: Hammonton Police Department , Hammonton NJ

Date Received: 04/16/12

Date given to service: 04/16/12

Carrier: FedEx UPS USPS Shipping Method: GRD 3DAY 2DAY
 NDA-PRI NDA-STD

Product: 6510 6810
 7110 8610
 7510 DT5000
 7410 Upper-half 9510
 7410 Whole

Serial Number: AR WF-0349

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: _____

Description: A B Plus Demo Screener Trade In

Accessories

110V A/C Adapter Regulator Printer Ribbon Printer Paper
 Mouthpieces 9510 Stylus 9510 Top Cover Carrying Case
 Dry Gas Other (please specify) _____

Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
8315075	PCB	1	N/C W
6808455	Fuel Cell	1	N/C W
MP CAL 71	Calibration	1	N/C W
MP Labor	Labor	.5	N/C W

Repair Notes: Replaced defective fuel cell number ARZK-0796 with new fuel cell number ARCM-1599.
While customer was conducting a linearity test for a 0.080% control test the instrument displayed a 6.464%.
Display cover was replaced due to being damage while being in service department. Cal W/QC AND OPS CHECK

Service Technician JS

Date: 04-25-12