

## RETURN AND REPAIR FORM

### Customer Information

B:

S:

Company Name: Ocean Twp P.D. , Waretown , N.J.

Date Received: 07/23/12

Date given to service: 07/23/12

Carrier:  FedEx  UPS  USPS Shipping Method:  GRD  3DAY  2DAY  
 NDA-PRI  NDA-STD

Product:  6510  6810 Serial Number: AR WF-0354  
 7110  8610 Printer Serial#: AR  
 7510  DT5000 Sim Serial#: DD  
 7410 Upper-half  9510 Probe Serial#: DD  
 7410 Whole

Warranty Expires: \_\_\_\_\_

Description:  A  B  Plus  Demo  Screener  Trade In

### Accessories

110V A/C Adapter  Regulator  Printer Ribbon  Printer Paper  
 Mouthpieces  9510 Stylus  9510 Top Cover  Carrying Case  
 Dry Gas  Other (please specify) \_\_\_\_\_

### Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
MP Cal 71	Calibration	1	NC/W
6808455	fuel cell	1	NC/W
Misc part	PCB 8315075	1	NC/W
MP Labor	Labor	.5	NC/W

Repair Notes: Replaced (PCB) because unit would not print from the external printer. Replaced fuel per-  
customers request for preventive maintenance, new fuel cell number # ARDD-1262.

CAL W/QC AND OPS CHECK

Service Technician JS

Date: 07-31-12