

# RETURN AND REPAIR FORM



## Customer Information

B:

S:

Company Name: Ocean Twp. P.D., Waretown, NJ

Date Received: 06/17/14

Date given to service: 06/17/14

Carrier:  FedEx  UPS  USPS Shipping Method:  GRD  3DAY  2DAY  
 NDA-PRI  NDA-STD

Product:  6510  6810  
 7110  8610  
 7510  DT5000  
 7410 Upper-half  9510  
 7410 Whole

Serial Number: AR WF-0354

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: Jan. 2016

Description:  A  B  Plus  Demo  Screener  Trade In

## Accessories

110V A/C Adapter  Regulator  Printer Ribbon  Printer Paper  
 Mouthpieces  9510 Stylus  9510 Top Cover  Carrying Case  
 Dry Gas  Other (please specify) \_\_\_\_\_

## Repair Information:

## Test#:

Part Number	Description	Qty	Total Cost
4414161	Calibration	1	NC/W
4414166	Labor	6	NC/W
6808455	Fuel Cell	1	NC/W
6808486	Spacer Plates	2	NC/W
12030	Cermetek Modem Chip	1	NC/W
6808374	Geared Engine (Motor)	1	NC/W
8100902	Flow Sensor	1	NC/W
8312820	Pump	1	NC/W

Repair Notes: Replaced defective fuel cell, spacer plates, Cermetek modem chip, geared engine, flow sensor, and pump. Fuel cell and spacer plates were changed per customer request for preventive maintenance. Customer reported multiple errors - modem, 31, 32, and 41. Changed modem chip, geared engine (motor), flow sensor, and pump corrected these issues. Also, blew out cuvette which had debris in it.

Fuel Cell = ARFD-1983

CAL W/QC AND OPS CHECK

Service Technician BC

Date: 07/11/14