

Return & Repair Form:

Customer Information:

Company Name: Bedminster B: 58861 S: _____
 Date Received: 5-21-10 Date Given to Service: _____
 Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY 3DY - Other

Product Information:

Product: 7410-7110-6510-6810-7510 Serial #: AR WF -0355
 9510- Drug Tester 5000
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR -
 Whole Instrument Top 1/2 Sim Ser #: _____
 Other _____ Probe Ser # DD P -
ACCESSORIES:
 110 V A/C Adapter Regulator Mag Card Rdr# _____
 Printer Paper Printer Ribbon Casio # _____
 Mouthpieces Carrying Case Dry Gas
 Other (Please Specify) _____ Warranty Exp. Date JAN. 2014

Repair Information:

Test #

Reason for Return: FC

Part Number	Description	Qty	Total Cost
<u>6808455</u>	<u>FC</u>	<u>1</u>	<u>N/C W</u>
<u>6808486</u>	<u>Plate</u>	<u>1</u>	<u>N/C W</u>
<u>MP CAL 71</u>	<u>CAL</u>	<u>1</u>	<u>N/C W</u>
<u>MP Labor</u>	<u>Labor</u>	<u>.5</u>	<u>N/C W</u>

Repair Notes:

NEW FC # ARAN-1039
CAL W/DC ; OPS CHECK

Service Technician: JF

Date: 05-24-2010