

Return & Repair Form

Customer Information:

CUSTOMER West Amwell Police Cust # 100518
 Date of Return: 12/4/2007 Carrier: FX UPS DHL USPS Method: GRD NDA 2DY 3DY
 OTHER: SAVER

Product Information:

Product: 7110 7410 6510 6810 Serial # AR WF - 0360
 Description: A B PLUS SCREENER _ Printer Serial # AR -
 Whole Inst. Top Half Simulator Serial # _____
 OTHER: _____ Probe Serial # DD P -
 Demo Unit Returned to stock on: _____
 Accessories: (check all that apply):
 110VAC Adapter Regulator Mag Card Rdr # _____
 Printer Paper Printer Ribbon Casio # _____
 Mouthpieces Carrying Case Dry Gas
 Other (specify): _____
 Warranty Expiration Date: _____

Repair Information:

Test # _____

Reason for Return: SPACE ON IR RESULT

Part #	Description	Qty	Total
<u>MPCAL 71</u>	<u>CAL</u>	<u>1</u>	<u>WARR N/C</u>
<u>MPLABOR</u>	<u>LABOR</u>	<u>.5</u>	<u>WARR N/C</u>

Repair Notes: REPLACED IR DETECTOR AS INSTRUCTED BY LÜBECK,
NON STOCK PN. 680945B N/C - FC WAS NEW IN '06 /
CAL W/ QC & OPS CHECKS

Technician: [Signature]

Date: 12-10-07