

# RETURN AND REPAIR FORM



## Customer Information

B:

S:

Company Name: Delaware Twp. P.D., Stockton, NJ

Date Received: 09/28/2016

Date given to service: 09/28/2016

Carrier:  FedEx  UPS  USPS      Shipping Method:  GRD  3DAY  2DAY  
 NDA-PRI  NDA-STD

Product:  6510  6810      Serial Number: AR WF-0361  
 7110  8610      Printer Serial#: AR  
 7510  DT5000      Sim Serial#: DD  
 7410 Upper-half  9510      Probe Serial#: DD  
 7410 Whole

Warranty Expires: \_\_\_\_\_

Description:  A  B  Plus  Demo  Screener  Trade In

## Accessories

110V A/C Adapter       Regulator       Printer Ribbon       Printer Paper  
 Mouthpieces       9510 Stylus       9510 Top Cover       Carrying Case  
 Dry Gas       Other (please specify) \_\_\_\_\_

## Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
4414161	Calibration	1	
4414166	Labor	1	
6808455	Fuel Cell	1	
6808486	Spacer Plates	2	

Repair Notes: Replaced defective fuel cell, c156, c167 and spacer plates.

C156 and C167 were changed due to modem issues.

Fuel Cell = ARJE-1885

CAL W/QC AND OPS CHECK

Service Technician BC

Date: 10/28/2016