

# Return & Repair Form:

## Customer Information:

Company Name: Phillipsburg PD B: \_\_\_\_\_ S: \_\_\_\_\_  
 Date Received: 10/25/2010 Date Given to Service: \_\_\_\_\_  
 Carrier: FedEx - UPS - DHL - USPS Method: ~~CRD~~ - ~~NDA~~ - 2DY - 3DY - Other

*ZDRY  
PER  
JOE*

## Product Information:

Product: 7410-7110-6510-6810-7510  
 9510- Drug Tester 5000  
 Description: A - B - Plus - Screener - Demo  
 Serial #: ARWF-0365  
 Printer Ser #: AR  
 Whole Instrument \_\_\_\_\_ Top 1/2 \_\_\_\_\_  
 Sim Ser #: \_\_\_\_\_  
 Other \_\_\_\_\_ Probe Ser # DD P  
**ACCESSORIES:**  
 110 V A/C Adapter       Regulator       Mag Card Rdr# \_\_\_\_\_  
 Printer Paper       Printer Ribbon       Casio # \_\_\_\_\_  
 Mouthpieces       Carrying Case       Dry Gas  
 Other (Please Specify) \_\_\_\_\_ Warranty Exp. Date \_\_\_\_\_

## Repair Information:

Test #

Reason for Return: \_\_\_\_\_

Part Number	Description	Qty	Total Cost
<u>MP Cal 71</u>	<u>Cal</u>	<u>1</u>	<u>n/c w</u>
<u>MP Labor</u>	<u>Labor</u>	<u>.5</u>	<u>n/c w</u>

### Repair Notes:

SCREW IN PRE-WARMING TRACK NOT SEATED - CORRECTED  
Cal w/ DC; OPS check

Service Technician: [Signature]

Date: 10/25/2010