

Return & Repair Form

Customer Information:

CUSTOMER LACEY TWP POLICE Cust # _____
 Date of Return: 1/23/2008 Carrier: EX UPS DHL USPS Method: GRD (NDA) 2DY 3DY
 OTHER: STUD

Product Information:

Product : (7110) 7410 6510 6810 Serial # AR WF - 0366
 Description: A B PLUS SCREENER Printer Serial # AR -
 Whole Inst. Top Half Simulator Serial # _____
 OTHER : _____ Probe Serial # DD P -

Demo Unit Returned to stock on: _____

Accessories: (check all that apply):

- 110VAC Adapter Regulator Mag Card Rdr # _____
- Printer Paper Printer Ribbon Casio # _____
- Mouthpieces Carrying Case Dry Gas
- Other (specify): _____

Warranty Expiration Date: _____

Repair Information:

Test # _____

Reason for Return: HIGH IR RESULTS

Part #	Description	Qty	Total
<u>MPCAL71</u>	<u>CAL</u>	<u>1</u>	<u>N/C WARR</u>
<u>MP LABOR</u>	<u>LABOR</u>	<u>.5</u>	<u>N/C WARR</u>

Repair Notes: REPLACED IR DETECTOR PN 6809458 (NON-STOCK PART)
CAL W/ QC & OPS CHECKS

Technician: [Signature]

Date: 01-24-08