

Dräger safety

Return & Repair Form

Customer Information:

CUSTOMER Venture City Police Cust # 57777
 Date of Return: 10/30/2007 Carrier: FX UPS DHL USPS Method: GRD NDA 2DY 3DY
 OTHER: _____

Product Information:

Product : 7110 7410 6510 6810 Serial # AR WF - 0368
 Description: A B PLUS SCREENER _ Printer Serial # AR -
 Whole Inst. Top Half Simulator Serial # _____
 OTHER : _____ Probe Serial # DD P -
 Demo Unit Returned to stock on: _____
 Accessories: (check all that apply):
 110VAC Adapter Regulator Mag Card Rdr # _____
 Printer Paper Printer Ribbon Casio # _____
 Mouthpieces Carrying Case Dry Gas
 Other (specify): _____
 Warranty Expiration Date: _____

Repair Information:

Test # _____

Reason for Return: _____

Part #	Description	Qty	Total
<u>6808455</u>	<u>FC</u>	<u>1</u>	<u>WARR N/C</u>
<u>MPCAL 71</u>	<u>CAL</u>	<u>1</u>	<u>WARR N/C</u>
<u>MPLABOR</u>	<u>LABOR</u>	<u>1.0</u>	<u>WARR N/C</u>

Repair Notes: REPLACED IR DETECTOR PN 6809458 (NON-STOCK)
PER LÜBECK INSTRUCTIONS - REPLACED FC AS P.M. / CAL W/ QC
AND OPS CHECKS

Technician: [Signature]

Date: 11-13-07