

## Return & Repair Form

### Customer Information:

CUSTOMER BERNARDS TWP P.D. Cust # \_\_\_\_\_  
 Date of Return: 09/21/2007 Carrier: FX UPS DHL USPS Method: GRD NDA 2DY 3DY  
 OTHER: STND O/N

### Product Information:

Product #: (7110) 7410 6510 6810 Serial # AR WF - 0371  
 Description: A B PLUS SCREENER \_ Printer Serial # AR -  
 Whole Inst. Top Half Simulator Serial # \_\_\_\_\_  
 OTHER: \_\_\_\_\_ Probe Serial # DD P -  
 Demo Unit  Returned to stock on: \_\_\_\_\_  
 Accessories: (check all that apply):  
 110VAC Adapter  Regulator  Mag Card Rdr # \_\_\_\_\_  
 Printer Paper  Printer Ribbon  Casio # \_\_\_\_\_  
 Mouthpieces  Carrying Case  Dry Gas  
 Other (specify): \_\_\_\_\_  
 Warranty Expiration Date: \_\_\_\_\_

### Repair Information:

Test # \_\_\_\_\_

Reason for Return: IR RUNS HIGH

| Part #          | Description     | Qty      | Total          |
|-----------------|-----------------|----------|----------------|
| <u>MPCAL 71</u> | <u>7110 CAL</u> | <u>1</u> | <u>WARRANT</u> |
|                 |                 |          |                |
|                 |                 |          |                |
|                 |                 |          |                |
|                 |                 |          |                |

Repair Notes: PERFORMED 50 STND-CHECKS-ALL OK & SHOW STANDARD SOLUTION DEFECTION / INSPECTED CUVEITE MIRROR = OK & CLEAR / FC = ~~WQ~~ -LEADS LOOK OK - PROBLEM NOT REPLICATED CAL, QC, OPS CHECKS

Technician: [Signature]

Date: 09.28.07