

Return & Repair Form

Customer Information:

CUSTOMER Bennards Twp Police Cust # S: 58728 B: 58726  
Date of Return: 03/12/2008 Carrier: FX UPS DHL USPS Method: GRD NDA 2DY 3DY  
OTHER: Standard

Product Information:

Product : 7110 7410 6510 6810 Serial # AR WF - 0371  
Description: A B PLUS SCREENER Printer Serial # AR -  
Whole Inst. Top Half Simulator Serial # \_\_\_\_\_  
OTHER : \_\_\_\_\_ Probe Serial # DD P -

Demo Unit  Returned to stock on: \_\_\_\_\_

Accessories: (check all that apply):

- 110VAC Adapter
- Regulator
- Mag Card Rdr # \_\_\_\_\_
- Printer Paper
- Printer Ribbon
- Casio # \_\_\_\_\_
- Mouthpieces
- Carrying Case
- Dry Gas
- Other (specify): \_\_\_\_\_

Warranty Expiration Date: \_\_\_\_\_

Repair Information:

Test # \_\_\_\_\_

Reason for Return: IR EC?

Part #	Description	Qty	Total
<u>6808455</u>	<u>FC</u>	<u>1</u>	<u>N/C WARR</u>
<u>MPMISC</u>	<u>680945B IR DETECTOR (NON-STOCK)</u>	<u>1</u>	<u>N/C WARR</u>
<u>MPCAL 71</u>	<u>7110 CIR</u>	<u>1</u>	<u>N/C WARR</u>
<u>MPCABOR</u>	<u>LABOR</u>	<u>.5</u>	<u>N/C WARR</u>

Repair Notes: REPLACED FC - REPLACED IR DETECTOR BASED ON DATA PROVIDED - CAL W/ QC OPS CHECKS

Technician: [Signature]

Date: 03-17-08