

Return & Repair Form:

59070

Customer Information:

Company Name: Phetcong PD B: _____ S: _____
 Date Received: 11/12/2010 Date Given to Service: _____
 Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY Other _____

Product Information:

Product: 7410-7110-6510-6810-7510
 9510- Drug Tester 5000
 Description: A - B - Plus - Screener - Demo
 Serial #: AR WF-0373
 Printer Ser #: AR _____
 Whole Instrument _____ Top 1/2 _____ Sim Ser #: _____
 Other _____ Probe Ser # DD _____ P _____
ACCESSORIES:
 110 V A/C Adapter Regulator Mag Card Rdr# _____
 Printer Paper Printer Ribbon Casio # _____
 Mouthpieces Carrying Case Dry Gas
 Other (Please Specify) _____ Warranty Exp. Date _____

Repair Information:

Test #

Reason for Return:

Part Number	Description	Qty	Total Cost
<u>8315075</u>	<u>PLIB</u>	<u>1</u>	<u>N/CW</u>
<u>6808455</u>	<u>FC</u>	<u>1</u>	<u>N/CW</u>
<u>6808486</u>	<u>Plate</u>	<u>1</u>	<u>N/CW</u>
<u>MP CAL 71</u>	<u>Cal</u>	<u>1</u>	<u>N/CW</u>
<u>MP Labor</u>	<u>Labor</u>	<u>1</u>	<u>N/CW</u>

Repair Notes:

FC# AR13E-1127, Install New PLIB

Cal w/ QC & ops check

Service Technician: 

Date: 12-1-2010