

Return & Repair Form

Customer Information:

CUSTOMER BERKELEY TWP P.D. Cust # _____
 Date of Return: 09/21/2007 Carrier: UPS DHL USPS Method: GRD NDA 2DY 3DY
 OTHER: PRIORITY O/N

Product Information:

Product 7110 7410 6510 6810 Serial # AR WF - 0374
 Description: A B PLUS SCREENER _ Printer Serial # AR -
 Whole Inst. Top Half Simulator Serial # _____
 Probe Serial # DD P -
 OTHER : _____
 Demo Unit Returned to stock on: _____
 Accessories: (check all that apply):
 110VAC Adapter Regulator Mag Card Rdr # _____
 Printer Paper Printer Ribbon Casio # _____
 Mouthpieces Carrying Case Dry Gas
 Other (specify): _____
 Warranty Expiration Date: _____

Repair Information:

Test # _____

Reason for Return: FAULTY DISPLAY

Part #	Description	Qty	Total
<u>8312928</u>	<u>DISPLAY</u>	<u>1</u>	<u>WARR</u>
<u>6808455</u>	<u>FE</u>	<u>1</u>	<u>WARR</u>
<u>MPCAL71</u>	<u>7110 CAL</u>	<u>1</u>	<u>WARR</u>
<u>MPLLABOR</u>	<u>LABOR</u>	<u>.5</u>	<u>WARR</u>

Repair Notes: EC IS WE-INSPECTED/ACID LEAK/REPLACED
REPLACED DISPLAY - CAL, QC, OPS CHECKS

Technician: [Signature]

Date: 09-27-07