

Return & Repair Form

Customer Information:

Company Name: Lucillette PD B: _____ S: _____
 Date Received: 7-28-08 Date Given to Service: _____
 Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

Product Information:

Product: 7410 - 7110 - 6510 - 6810 Serial #: AR WF - 0375
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR -
 Whole Instrument Top 1/2 Sim Ser #: _____
 Other _____ Probe Ser # DD _____ P -
ACCESSORIES:
 110 V A/C Adapter Regulator Mag Card Rdr# _____
 Printer Paper Printer Ribbon Casio # _____
 Mouthpieces Carrying Case Dry Gas
 Other (Please Specify) _____ Warranty Exp. Date _____

Repair Information:

Reason for Return:		Test #	
PROBE CIRC. + FC + AMB AIR CHECK PROBS			
Part Number	Description	Qty	Total Cost
<u>6808455</u>	<u>FC</u>	<u>1</u>	<u>N/C WARR</u>
<u>NIPCAL 71</u>	<u>CAL</u>	<u>1</u>	<u>N/C WARR</u>
<u>MPLABOR</u>		<u>0.5</u>	<u>N/C WARR</u>

Repair Notes:

REPLACED FC / RE-CAL PROBE CIRC. / CAL W/ QC & OPS CHECKS

AMB AIR CHECK PROBLEM APPEARS TO BE ENVIRONMENTAL PER DATA (NOT GOOD ON POST AIR BLANKS)

Service Technician: [Signature]

Date: 07-29-08