

Return & Repair Form

Customer Information:

CUSTOMER LAKEHURST POLICE Cust # _____
Date of Return: 12/21/06 Carrier: FED EX Method: Priority

Product Information:

Product : 7110 7410 6510 6810 Serial # ARWF-0376

Description: A B PLUS SCREENER Printer Serial # _____

Whole Inst. Top Half Simulator Serial # _____

OTHER : _____ Probe Serial # _____

Demo Unit Returned to stock on: _____

Accessories: (check all that apply):

- | | | |
|---|---|---|
| <input type="checkbox"/> 110VAC Adapter | <input type="checkbox"/> Regulator | <input type="checkbox"/> Mag Card Rdr # _____ |
| <input type="checkbox"/> Printer Paper | <input type="checkbox"/> Printer Ribbon | <input type="checkbox"/> Casio # _____ |
| <input type="checkbox"/> Mouthpieces | <input type="checkbox"/> Carrying Case | <input type="checkbox"/> Dry Gas |
| <input type="checkbox"/> Other (specify): _____ | | |

Warranty Expiration Date: _____

Repair Information:

Test # _____

Reason for Return: Unit failed in-Test.

Part #	Description	Qty	Total
<u>MPCAL71</u>	<u>7110 CAL</u>	<u>1</u>	

Repair Notes: Unit verified to be running high. Re-Cald unit and ran in-TEST. Ops check + O.C.

Technician: CRD

Date: 12/22/06