

## Return & Repair Form

### Customer Information:

CUSTOMER Brandenburg TWP PD Cust # \_\_\_\_\_  
Date of Return: 03/24/2008 Carrier:  UPS  DHL  USPS Method: GRD  NDA  2DY  3DY  
OTHER: Standard

### Product Information:

Product : 110 7410 6510 6810 Serial # AR WF - 0377  
Description: A B PLUS SCREENER Printer Serial # AR -  
Whole Inst. Top Half Simulator Serial # \_\_\_\_\_  
OTHER : \_\_\_\_\_ Probe Serial # DD P -  
Demo Unit  Returned to stock on: \_\_\_\_\_  
Accessories: (check all that apply):  
 110VAC Adapter  Regulator  Mag Card Rdr # \_\_\_\_\_  
 Printer Paper  Printer Ribbon  Casio # \_\_\_\_\_  
 Mouthpieces  Carrying Case  Dry Gas  
 Other (specify): \_\_\_\_\_  
Warranty Expiration Date: \_\_\_\_\_

### Repair Information:

Test # \_\_\_\_\_

Reason for Return: FC + IR RESULTS

| Part #          | Description                                    | Qty       | Total           |
|-----------------|------------------------------------------------|-----------|-----------------|
| <u>6808455</u>  | <u>FC</u>                                      | <u>1</u>  | <u>N/C WARR</u> |
| <del>6809</del> |                                                |           |                 |
| <u>MPMISC</u>   | <u>(PN 680945B) IR DETECT. PCB (NON STOCK)</u> | <u>1</u>  | <u>N/C WARR</u> |
| <u>MPCAL 71</u> | <u>7110 CAL</u>                                | <u>1</u>  | <u>N/C WARR</u> |
| <u>MP LABOR</u> | <u>LABOR</u>                                   | <u>.5</u> | <u>N/C WARR</u> |

Repair Notes: REPLACE FC + IR DETECT. PCB  
CAL w/ QC & OPS CHECKS

Technician: [Signature]

Date: 03.25.08