

Return & Repair Form

Customer Information:

Company Name: Branchburg Twp B: _____ S: _____
 Date Received: 9-10-08 Date Given to Service: _____
 Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

Product Information:

Product: 7410 - 7110 - 6510 - 6810 Serial #: AR WF 0377
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR -
 Whole Instrument Top 1/2 Sim Ser #: _____
 Other _____ Probe Ser # DD P -
ACCESSORIES:
 110 V A/C Adapter Regulator Mag Card Rdr# _____
 Printer Paper Printer Ribbon Casio # _____
 Mouthpieces Carrying Case Dry Gas
 Other (Please Specify) _____ Warranty Exp. Date _____

Repair Information:

Test #

Reason for Return:			
Part Number	Description	Qty	Total Cost
<u>6807500</u>	<u>CUVETTE</u>	<u>1</u>	<u>NIC WARR</u>
<u>MPCAL71</u>	<u>CAL</u>	<u>1</u>	<u>NIC WARR</u>
<u>MPLABOR</u>		<u>1 HR</u>	

Repair Notes:

REPLACED CUVETTE (IR DETECTOR WAS REPLACED 3-25-08 AND PROBLEM WITH IR STILL HAPPENS.)

CAL w/ QC & OPS CHECKS

Service Technician: _____

Date: 09/11/08