

# RETURN AND REPAIR FORM



## Customer Information

B:

S:

Company Name: Belvidere P.D., Belvidere, NJ

Date Received: 02/28/12 Date given to service: 02/28/12

Carrier:  FedEx  UPS  USPS Shipping Method:  GRD  3DAY  2DAY  
 NDA-PRI  NDA-STD

Product:  6510  6810 Serial Number: AR WF-0381  
 7110  8610 Printer Serial#: AR  
 7510  DT5000 Sim Serial#: DD  
 7410 Upper-half  9510 Probe Serial#: DD  
 7410 Whole

Warranty Expires: \_\_\_\_\_

Description:  A  B  Plus  Demo  Screener  Trade In

## Accessories

110V A/C Adapter  Regulator  Printer Ribbon  Printer Paper  
 Mouthpieces  9510 Stylus  9510 Top Cover  Carrying Case  
 Dry Gas  Other (please specify) \_\_\_\_\_

## Repair Information:

## Test#:

Part Number	Description	Qty	Total Cost
<u>6808486</u>	<u>Plate</u>	<u>1</u>	<u>N/C W</u>
<u>6808455</u>	<u>Fuel cell</u>	<u>1</u>	<u>N/C W</u>
<u>mp cal 71</u>	<u>Calibration</u>	<u>1</u>	<u>N/C W</u>
<u>mp labor</u>	<u>labor</u>	<u>1.5</u>	<u>N/C W</u>

Repair Notes: Replaced Fuel cell # ARWE-0118 with Fuel cell # ARCK-0903.

cal w/QL, ops check

Service Technician: [Signature] Date: 3.9.12