

Return & Repair Form

Customer Information:

CUSTOMER WEST WINDSOR Twp PD Cust # _____

Date of Return: 4/19 Carrier: Fed Exp Method: PRIORITY

Product Information:

Product: 7110 7410 6510 6810 Serial # ARWF-0382

Description: A B PLUS SCREENER _____ Printer Serial # _____

Whole Inst. Top Half Simulator Serial # _____

OTHER: _____ Probe Serial # _____

Demo Unit Returned to stock on: _____

Accessories: (check all that apply):

- 110VAC Adapter
- Regulator
- Mag Card Rdr # _____
- Printer Paper
- Printer Ribbon
- Casio # _____
- Mouthpieces
- Carrying Case
- Dry Gas
- Other (specify): _____

Warranty Expiration Date: _____

Repair Information:

Test # _____

Reason for Return: Random high IR readings

| Part # | Description | Qty | Total |
|---------|-------------|-----|-------|
| MPCAL71 | 7110 CAL | 1 | N/C |
| | | | |
| | | | |
| | | | |
| | | | |
| | | | |

Repair Notes: Could not duplicate. Cleaned spot of mirror.
Ran 99 + STD-CHECK without any problems.
Cal ops check + O.C.

Technician: CAB

Date: 4/26/06