

# Dräger safety

Internal Use Only

## Return & Repair Form

### Customer Information:

CUSTOMER West Windsor Twp PD Cust # \_\_\_\_\_  
Date of Return: \_\_\_\_\_ Carrier: \_\_\_\_\_ Method: FedEx Express

### Product Information:

Product: 7110 7410 6510 6810 Serial # ARWF-0382  
Description: A B PLUS SCREENER \_\_\_\_\_ Printer Serial # \_\_\_\_\_  
Whole Inst. Top Half Simulator Serial # \_\_\_\_\_  
OTHER: \_\_\_\_\_ Probe Serial # \_\_\_\_\_  
Demo Unit  Returned to stock on: \_\_\_\_\_

### Accessories: (check all that apply):

- 110VAC Adapter  Regulator  Mag Card Rdr # \_\_\_\_\_  
 Printer Paper  Printer Ribbon  Casio # \_\_\_\_\_  
 Mouthpieces  Carrying Case  Dry Gas  
 Other (specify): \_\_\_\_\_

Warranty Expiration Date: \_\_\_\_\_

### Repair Information:

Test # \_\_\_\_\_

Reason for Return: Random High IR

Part #	Description	Qty	Total
<u>MPCAL71</u>	<u>7110 CAL</u>	<u>1</u>	

Repair Notes: Could not duplicate random high IR readings after  $\approx$  300 tests. Full cal ops check & O.C.

Technician: CRD

Date: 11/15/06