

Return & Repair Form

Customer Information:

CUSTOMER Montgomery TWSH P.D. Cust # 58592
 Date of Return: 12 7 / 2007 Carrier: UPS DHL USPS Method: GRD NDA 2DY 3DY
 OTHER: _____

Product Information:

Product: 7110 7410 6510 6810 Serial # ARWF-0384
 Description: A B PLUS SCREENER _ Printer Serial # AR -
 Whole Inst. Top Half Simulator Serial # _____
 OTHER: _____ Probe Serial # DD P -

Demo Unit Returned to stock on: _____

Accessories: (check all that apply):

- 110VAC Adapter
- Regulator
- Mag Card Rdr # _____
- Printer Paper
- Printer Ribbon
- Casio # _____
- Mouthpieces
- Carrying Case
- Dry Gas
- Other (specify): _____

Warranty Expiration Date: _____

Repair Information:

Test # _____

Reason for Return: ~~FC~~ LOW FC RESULTS

Part #	Description	Qty	Total
<u>680845</u>	<u>FC</u>	<u>1</u>	<u>WARR N/C</u>
<u>MPCAL 71</u>	<u>CAL</u>	<u>1</u>	<u>WARR N/C</u>
<u>MPLABOR</u>	<u>LABOR</u>	<u>.5</u>	<u>WARR N/C</u>

Repair Notes: REPLACE FC / CAL W QC & OPS CHECKS

Technician: [Signature]

Date: 12-10-07