

Return & Repair Form:

Customer Information:

Company Name: Montgomery Trust B: _____ S: _____
 Date Received: 04/28/2009 Date Given to Service: 04/28/2009
 Carrier: RedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

Product Information:

Product: 7410 7110 - 6510 - 6810 Serial #: AR WF-0384
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR _____
 Whole Instrument Top 1/2 Simi Ser #: _____
 Other _____ Probe Ser # DD _____ P _____
ACCESSORIES:
 110 V A/C Adapter Regulator Mag Card Rdr# _____
 Printer Paper Printer Ribbon Casio # _____
 Mouthpieces Carrying Case Dry Gas
 Other (Please Specify) _____ Warranty Exp. Date _____

Repair Information:

FC

Test #

Reason for Return: _____

Part Number	Description	Qty	Total Cost
<u>0808455</u>	<u>FC</u>	<u>1</u>	<u>N/C WARR</u>
<u>12003</u>	<u>HOSE BARB</u>	<u>1</u>	<u>N/C WARR</u>
<u>MP CAL 71</u>	<u>CAL</u>	<u>1</u>	<u>N/C WARR</u>
<u>MP Labor</u>	<u>Labor</u>	<u>.5</u>	<u>N/C WARR</u>

Repair Notes:

Replace FC / HOSE BARB
Cal w/QC & Ops Check

Service Technician: [Signature]

Date: 04/29/2009