

# Return & Repair Form:

## Customer Information:

Company Name: Northvale Police Dept. B: \_\_\_\_\_ S: \_\_\_\_\_

Date Received: 2-5-10 Date Given to Service: \_\_\_\_\_

Carrier: FedEx UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY Other

## Product Information:

Product: 7410-7110-6510-6810-7510  
9510- Drug Tester 5000

Serial #: AR WF-0387

Description: A - B - Plus - Screener - Demo

Printer Ser #: AR -

Whole Instrument

Top 1/2

Sim Ser #: \_\_\_\_\_

Other \_\_\_\_\_

Probe Ser # DD P -

## ACCESSORIES:

110 V A/C Adapter

Regulator

Mag Card Rdr# \_\_\_\_\_

Printer Paper

Printer Ribbon

Casio # \_\_\_\_\_

Mouthpieces

Carrying Case

Dry Gas

Other (Please Specify) \_\_\_\_\_

Warranty Exp. Date \_\_\_\_\_

## Repair Information:

Test #

Reason for Return: FC

Part Number	Description	Qty	Total Cost
<u>6808455</u>	<u>FC</u>	<u>1</u>	<u>N/C WGM</u>
<u>MP CAL 71</u>	<u>CAL</u>	<u>1</u>	<u>N/C WGM</u>
<u>MP Labor</u>	<u>Labor</u>	<u>.5</u>	<u>N/C WGM</u>

Repair Notes:

Replaced FC # ARWE-0071 with FC # ARAM-0759

Cal w/OC 1/3 Ops check

Service Technician: [Signature]

Date: 02/8/2010