

# Return & Repair Form:

## Customer Information:

Company Name: Lakewood Twp PA B: \_\_\_\_\_ S: \_\_\_\_\_  
 Date Received: 7-17-2009 Date Given to Service: \_\_\_\_\_  
 Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

## Product Information:

Product: 7410 7110 - 6510 - 6810 Serial #: AR WF - 0391  
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR -  
 Whole Instrument Top 1/2 Sim Ser #: \_\_\_\_\_  
 Other \_\_\_\_\_ Probe Ser # DD P -  
**ACCESSORIES:**  
 110 V A/C Adapter  Regulator  Mag Card Rdr# \_\_\_\_\_  
 Printer Paper  Printer Ribbon  Casio # \_\_\_\_\_  
 Mouthpieces  Carrying Case  Dry Gas  
 Other (Please Specify) \_\_\_\_\_ Warranty Exp. Date \_\_\_\_\_

## Repair Information:

Reason for Return: FC/CUVette Test #

Part Number	Description	Qty	Total Cost
<u>6809500</u>	<u>CUVette</u>	<u>1</u>	<u>N/C W/ WARR</u>
<u>6808486</u>	<u>Platen</u>	<u>1</u>	<u>N/C W/ WARR</u>
<u>6808455</u>	<u>FC</u>	<u>1</u>	<u>N/C W/ WARR</u>
<u>MP Cal 71</u>	<u>Cal</u>	<u>1</u>	<u>N/C W/ WARR</u>
<u>MP Labor</u>	<u>Labor</u>	<u>.5</u>	<u>N/C W/ WARR</u>

## Repair Notes:

Replace CUVette, FC  
Cal w/ Qc; Ops check

Service Technician: AS

Date: 07-22-2009