

# RETURN AND REPAIR FORM



## Customer Information

B: \_\_\_\_\_ S: \_\_\_\_\_

Company Name: Lakewood P.D., Lakewood, NJ

Date Received: 02/27/15

Date given to service: 02/27/15

Carrier:  FedEx  UPS  USPS      Shipping Method:  GRD  3DAY  2DAY  
 NDA-PRI  NDA-STD

Product:  6510  6810  
 7110  8610  
 7510  DT5000  
 7410 Upper-half  9510  
 7410 Whole

Serial Number: AR WF-0391

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: January 2011 (60128)

Description:  A  B  Plus  Demo  Screener  Trade In

## Accessories

110V A/C Adapter  Regulator  Printer Ribbon  Printer Paper  
 Mouthpieces  9510 Stylus  9510 Top Cover  Carrying Case  
 Dry Gas  Other (please specify) \_\_\_\_\_

## Repair Information:

Test#: \_\_\_\_\_

Part Number	Description	Qty	Total Cost
4414161	Calibration	1	
4414166	Labor	5	
6808455	Fuel Cell	1	
6808486	Spacer Plates	2	
8315075	Alcotest 7110 Motherboard	1	

Repair Notes: Replaced defective motherboard, fuel cell, and spacer plates.

Changed motherboard due to issues with the simulator temperature listing as OK instead of showing temperature.

Fuel Cell = ARFJ-0711

CAL W/QC AND OPS CHECK

Service Technician BC

Date: 04/02/15