Return & Repair Form:

Customer Information:

Company Name: Toms River B: _____ S: _____

Date Received: 1-7-09 Date Given to Service: _____

Carrier: FedEx – UPS – DHL – USPS **Method:** GRD – NDA – 2DY – 3DY - Other

Product Information: ~~Priority~~ Standard

Product: 7410 – 7110 – 6510 – 6810 Serial #: AR WF-0394

Description: A – B – Plus – Screener – Demo Printer Ser #: AR _____ - _____

Whole Instrument **Top ½** Sim Ser #: _____

Other _____

ACCESSORIES:

☐ 110 V A/C Adapter ☐ Regulator

☐ Printer Paper ☐ Printer Ribbon

☐ Mouthpieces ☐ Carrying Case

Other (Please Specify) _____

PER STEVE D.
PROBLEM MAY BE CAUSED BY ROOM/ENVIRONMENT- THAT THE 7110 IS IN-

CHEIF HAD BEEN ADVISED TO MOVE UNIT TO ANOTHER ROOM BUT REFUSED

Repair Information:

Reason for Return:

Part Number	Description	Qty	
6808455	FC	1	N/C WARR
6809458	IR DETECTOR	1	N/C WARR
MPCM 41	CM	1	N/C WARR
MPLABOR		.5	N/C WARR

Repair Notes:

BREATH HOSE HAD PLASTIC FROM MOUTHPIECE BAG BLOCKING BREATH HOSE PATH – REPLACED FC AS P.M. DUE TO TEST DATA (OLD CELL WAS YA) – REPLACED IR DETECTOR

CAL W/ QC & OPS CHECKS

Service Technician: [signature] Date: 01-08-09