

Return & Repair Form:

Customer Information:

Company Name: Toms River P.D. B: _____ S: _____
 Date Received: 1-25-10 Date Given to Service: _____
 Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

Product Information:

Product: 7410-7110-6510-6810-7510 Serial #: AR WF - 0394
 9510- Drug Tester 5000
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR -
 Whole Instrument Top 1/2 Sim Ser #: _____
 Other _____ Probe Ser # DD P -
ACCESSORIES:
 110 V A/C Adapter Regulator Mag Card Rdr# _____
 Printer Paper Printer Ribbon Casio # _____
 Mouthpieces Carrying Case Dry Gas
 Other (Please Specify) _____ Warranty Exp. Date _____

Repair Information:

Test #

Reason for Return:

Part Number	Description	Qty	Total Cost
<u>MP Cal 71</u>	<u>Cal</u>	<u>1</u>	<u>N/C WARR</u>
<u>Misc Part</u>	<u>Pressure sensor</u>	<u>1</u>	<u>N/C WARR</u>
<u>MP Cal</u>		<u>.5</u>	<u>N/C WARR</u>

Repair Notes:

Clean spot OFF CUvette.

Cal w/QC & Ops Check

Reclaimed Part Pressure sensor

Service Technician: JS

Date: 4/27/2010