

Return & Repair Form

Customer Information:

CUSTOMER SOMERS POINT CITY P.D. Cust # _____
 Date of Return: 12/14/2007 Carrier: FX UPS DHL USPS Method: GRD NDA 2DY 3DY
 OTHER: _____

Product Information:

Product : 7110 7410 6510 6810 Serial # ARWF - 0398
 Description: A B PLUS SCREENER _ Printer Serial # AR -
 Whole Inst. Top Half Simulator Serial # _____
 OTHER : _____ Probe Serial # DD P -
 Demo Unit Returned to stock on: _____
 Accessories: (check all that apply):
 110VAC Adapter Regulator Mag Card Rdr # _____
 Printer Paper Printer Ribbon Casio # _____
 Mouthpieces Carrying Case Dry Gas
 Other (specify): _____
 Warranty Expiration Date: _____

Repair Information:

Test # _____

Reason for Return: IR-FC
* REPLACED DUE TO DAMAGED CASE

Part #	Description	Qty	Total
<u>12008</u>	<u>BUMPER</u>	<u>2</u>	<u>* N/C TO CUSTOMER</u>
<u>6808455</u>	<u>FC</u>	<u>1</u>	<u>WARR N/C</u>
<u>6809458</u>	<u>IR DETECTOR</u>	<u>1</u>	<u>* WARR N/C</u>
<u>8314685</u>	<u>LOWER CASE</u>		<u>* N/C TO CUSTOMER</u>
<u>MPLCALZFI</u>	<u>CAL</u>	<u>1</u>	<u>WARR N/C</u>
<u>MPLLABOR</u>	<u>LABOR</u>	<u>1</u>	<u>* N/C TO CUSTOMER</u>

Repair Notes: LOWER CASE DAMAGED BY FED EX - REPLACE LOWER CASE w/ BUMPERS - REPLACED IR DETECTOR PER LUBBECK INSTRUCTIONS / REPLACE FC / CAL w/ QC SOPS CHECKS

Technician: [Signature]

Date: 12-17-07