

Return & Repair Form

Customer Information:

Company Name: Somerset Point PD B: 5599 S: Same
 Date Received: 6-9-08 Date Given to Service: 6-10-08
 Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY 3DY Other

Product Information:

Product: 7410 7110 - 6510 - 6810 Serial #: AR WF - 0398
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR -
 Whole Instrument Top 1/2 Sim Ser #: _____
 Other _____ Probe Ser # DD _____ P -
ACCESSORIES:
 110 V A/C Adapter Regulator Mag Card Rdr# _____
 Printer Paper Printer Ribbon Casio # _____
 Mouthpieces Carrying Case Dry Gas
 Other (Please Specify) _____ Warranty Exp. Date _____

Repair Information:

Test #

Reason for Return: IR Jump in Results

Part Number	Description	Qty	Total Cost
<u>MPMISC</u>	<u>IR DETECTOR (NON INV PN 680945B)</u>	<u>1</u>	<u>N/C WARR</u>
<u>MPMISC</u>	<u>IR FILTER NON INV.</u>	<u>1</u>	<u>N/C WARR</u>
<u>MPCAL 71</u>	<u>7110 CMC</u>	<u>1</u>	<u>N/C WARR</u>
<u>MPLABOR</u>		<u>.5</u>	<u>N/C WARR</u>

Repair Notes:

FC 15 YD INSTALLED 12.17.07 / SMALL SPOT ON IR FILTER - REPLACED
w/ NON INVENTORY PART // CAL W/ QC & OPS CHECKS
REPLACED IR DETECT PCB AS PREVENT. MEASURE

TESTED UNIT (50 STD CHECKS) PRIOR TO ANY REPAIR / ALL TESTS
GOOD! SEE DATA.

Service Technician: [Signature]

Date: 06.11.08