

Return & Repair Form:

**Customer Information:**

Company Name: Somer Point PD B: \_\_\_\_\_ S: \_\_\_\_\_  
 Date Received: 1/21/2011 Date Given to Service: \_\_\_\_\_  
 Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

**Product Information:**

Product: 7410-7110-6510-6810-7510 Serial #: ARWF-0398  
 9510- Drug Tester 5000  
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR  
 Whole Instrument Top 1/2 Sim Ser #: \_\_\_\_\_  
 Other \_\_\_\_\_ Probe Ser # DD P -

**ACCESSORIES:**

110 V A/C Adapter  Regulator  Mag Card Rdr# \_\_\_\_\_  
 Printer Paper  Printer Ribbon  Casio # \_\_\_\_\_  
 Mouthpieces  Carrying Case  Dry Gas  
 Other (Please Specify) \_\_\_\_\_ Warranty Exp. Date \_\_\_\_\_

**Repair Information:**

Test #

Reason for Return: FC

Part Number	Description	Qty	Total Cost
<u>6808435</u>	<u>FC</u>	<u>1</u>	<u>N/C</u>
<u>MP Cal 71</u>	<u>Cal</u>	<u>1</u>	<u>N/C</u>
<u>MP Labor</u>	<u>Labor</u>	<u>.5</u>	<u>N/C</u>

Repair Notes:

FC ~~7~~ AR 13L-1044  
Cal w/DC & OPS check

Service Technician: [Signature]

Date: 1/25/2011