



RETURN AND REPAIR FORM

Customer Information

B:

S:

Company Name: Aberdeen Twp P.D., Aberdeen NJ

Date Received: 03/02/12 Date given to service: 03/02/12

Carrier: FedEx UPS USPS Shipping Method: GRD 3DAY 2DAY
 NDA-PRI NDA-STD

Product: 6510 6810
 7110 8610
 7510 DT5000
 7410 Upper-half 9510
 7410 Whole

Serial Number: AR WF-0406

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: _____

Description: A B Plus Demo Screener Trade In

Accessories

110V A/C Adapter Regulator Printer Ribbon Printer Paper
 Mouthpieces 9510 Stylus 9510 Top Cover Carrying Case
 Dry Gas Other (please specify) _____

Repair Information:

Test#:

| Part Number | Description | Qty | Total Cost |
|------------------|--------------------|-----------|--------------|
| <u>6808455</u> | <u>Fuel cell</u> | <u>1</u> | <u>n/c w</u> |
| <u>mp Cal 71</u> | <u>Calibration</u> | <u>1</u> | <u>n/c w</u> |
| <u>mp Labor</u> | <u>Labor</u> | <u>.5</u> | <u>n/c w</u> |

Repair Notes: Replaced busted Fuel cell AR-2292 with Fuel cell ARCK-08795 ARCK-0879.

Cal w/DC; ops check