

Dräger safety

Return & Repair Form

Customer Information:

CUSTOMER Police Cust # _____
 Date of Return: 9/1/06 Carrier: FED EX Method: Exp-Saver

Product Information:

Product: 7110 7410 6510 6810 Serial # ARWJ-0006
 Description: A B PLUS SCREENER _ Printer Serial # _____
 Whole Inst. Top Half Simulator Serial # _____
 OTHER: _____ Probe Serial # _____
 Demo Unit Returned to stock on: _____
Accessories: (check all that apply):
 110VAC Adapter Regulator Mag Card Rdr # _____
 Printer Paper Printer Ribbon Casio # _____
 Mouthpieces Carrying Case Dry Gas
 Other (specify): _____
 Warranty Expiration Date: _____

Repair Information:

Test # _____

Reason for Return: LOW EC Readings on SOLW CHANGE only

Part #	Description	Qty	Total
<u>6808455</u>	<u>Fuel Cell</u>	<u>1</u>	
<u>MPCAL71</u>	<u>7110 CAL</u>	<u>1</u>	
<u>MPLABOR</u>	<u>LABOR</u>	<u>3</u>	

Repair Notes: EC readings confirmed low on 2nd + 3rd tests during SOLW-CHANGE ONLY. Found that FIC cause this. Replaced FIC. Cal ops check + Q.C.

Technician: CPD

Date: 9/1/06