

# Return & Repair Form:

## Customer Information:

Company Name: Watchung PD B: \_\_\_\_\_ S: \_\_\_\_\_  
 Date Received: 6-29-11 Date Given to Service: 6-29-11  
 Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

## Product Information:

Product: ~~7410-7110-6510-6810-7510~~ 9510- Drug Tester 5000 Serial #: ARWJ - 0013  
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR -  
 Whole Instrument Top ½ Sim Ser #: \_\_\_\_\_  
 Other \_\_\_\_\_ Probe Ser # DD P -

## ACCESSORIES:

110 V A/C Adapter  Regulator  Mag Card Rdr# \_\_\_\_\_  
 Printer Paper  Printer Ribbon  Casio # \_\_\_\_\_  
 Mouthpieces  Carrying Case  Dry Gas  
 Other (Please Specify) \_\_\_\_\_ Warranty Exp. Date \_\_\_\_\_

## Repair Information:

Test #

Reason for Return: \_\_\_\_\_

Part Number	Description	Qty	Total Cost
<u>6808455</u>	<u>FC</u>	<u>1</u>	<u>N/C W</u>
<u>mp Cal 71</u>	<u>Cal</u>	<u>1</u>	<u>N/C W</u>
<u>mp Labor</u>	<u>Labor</u>	<u>.5</u>	<u>N/C W</u>

## Repair Notes:

ARCA-0689  
Cal w/DC, ops check

Service Technician: [Signature]

Date: 7-5-11