

Return & Repair Form:

Customer Information:

Company Name: Point Pleasant Boro B: _____ S: _____

Date Received: 8-5-09 Date Given to Service: _____

Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

Product Information:

Product: 7410 - 7110 - 6510 - 6810

Serial #: AR WJ - 0014

Description: A - B - Plus - Screener - Demo

Printer Ser #: AR -

Whole Instrument Top 1/2

Sim Ser #: _____

Other _____

Probe Ser # DD P -

ACCESSORIES:

- 110 V A/C Adapter
- Regulator
- Mag Card Rdr# _____
- Printer Paper
- Printer Ribbon
- Casio # _____
- Mouthpieces
- Carrying Case
- Dry Gas

Other (Please Specify) _____ Warranty Exp. Date _____

Repair Information:

PCB

Test #

Reason for Return:

Part Number	Description	Qty	Total Cost
<u>8315075</u>	<u>PCB</u>	<u>1</u>	<u>N/C Labor</u>
<u>MP Cal 71</u>	<u>Cal</u>	<u>1</u>	<u>N/C Labor</u>
<u>MP Labor</u>	<u>Labor</u>	<u>1.0</u>	<u>N/C Labor</u>

Repair Notes:

Replace PCB
Cal w/ Qc & Ops Check

Service Technician: [Signature]

Date: 08-06-09