

RETURN AND REPAIR FORM



Customer Information

B:

S:

Company Name: Middletown Twp. P.D., Middletown, NJ

Date Received: 07/07/15

Date given to service: 07/07/15

Carrier: FedEx UPS USPS Shipping Method: GRD 3DAY 2DAY
 NDA-PRI NDA-STD

Product: 6510 6810
 7110 8610
 7510 DT5000
 7410 Upper-half 9510
 7410 Whole

Serial Number: AR WJ-0015

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: November 2016 (61153)

Description: A B Plus Demo Screener Trade In

Accessories

110V A/C Adapter Regulator Printer Ribbon Printer Paper
 Mouthpieces 9510 Stylus 9510 Top Cover Carrying Case
 Dry Gas Other (please specify) _____

Repair Information:

Test#:

| Part Number | Description | Qty | Total Cost |
|-------------|---------------------------|-----|------------|
| 4414161 | Calibration | 1 | NC/W |
| 4414166 | Labor | 4 | NC/W |
| 8315075 | Alcotest 7110 Motherboard | 1 | NC/W |
| | | | |
| | | | |
| | | | |
| | | | |

Repair Notes: Replaced defective motherboard.

Unit was displaying improper time and changing motherboard resolved this issue.

CAL W/QC AND OPS CHECK

Service Technician BC

Date: 07/14/15