

Return & Repair Form:

Customer Information:

Company Name: Fred Astor PD B: _____ S: _____
 Date Received: 1-13-09 Date Given to Service: _____
 Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

Product Information:

Product: 7410 7110 - 6510 - 6810 Serial #: AR WJ - 0016
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR
 Whole Instrument Top 1/2 Sim Ser #: _____
 Other _____ Probe Ser # DD P -

ACCESSORIES:

110 V A/C Adapter Regulator Mag Card Rdr# _____
 Printer Paper Printer Ribbon Casio # _____
 Mouthpieces Carrying Case Dry Gas
 Other (Please Specify) _____ Warranty Exp. Date _____

Repair Information:

Reason for Return: SOFT UNIT LOCKS UP Test # _____

| Part Number | Description | Qty | Total Cost |
|-----------------|-------------|-------------|-----------------|
| <u>8315075</u> | <u>PCB</u> | <u>1</u> | <u>N/C WARR</u> |
| <u>MPCAL 71</u> | <u>CAL</u> | <u>1</u> | <u>N/C WARR</u> |
| | | | |
| | | | |
| | | | |
| <u>MPLABOR</u> | | <u>1 HR</u> | <u>N/C WARR</u> |

Repair Notes:

PROBLEM REPLICATED / REPLACED PCB / FC IS ZIC - WAS
INSTALLED ON 11-25-08
CAL W/ QC & OPS CHECKS

Service Technician: [Signature]

Date: 01-19-09