

Return & Repair Form:

Customer Information:

Company Name: BEECHWOOD POLICE DEPT. B: _____ S: _____

Date Received: 12-04-08 Date Given to Service: 12-04-08

Carrier: FedEx UPS DHL USPS Method: GRD NDA 2DY 3DY Other

Product Information:

Product: 7410- 7110 6510 - 6810 Serial #: AR WJ - 0017

Description: A - B - Plus - Screener - Demo Printer Ser #: AR -

Whole Instrument Top 1/2 Sim Ser #: _____

Other _____ Probe Ser # DD P -

ACCESSORIES:

110 V A/C Adapter Regulator Mag Card Rdr# _____

Printer Paper Printer Ribbon Casio # _____

Mouthpieces Carrying Case Dry Gas

Other (Please Specify) _____ Warranty Exp. Date _____

Repair Information:

Test #

Reason for Return:

Part Number	Description	Qty	Total Cost
<u>6809500</u>	<u>CUVETTE</u>	<u>1</u>	<u>N/C WORK</u>
<u>MPCACT1</u>	<u>CR</u>	<u>1</u>	<u>N/C WORK</u>
<u>MPLABOR</u>		<u>.5</u>	<u>N/C WORK</u>

Repair Notes:

IR DETECTOR WAS REPLACED 11-2007 - UNIT STILL HAS RANDOM PROBLEMS
REPLACED CUVETTE - FC WAS INSTALLED 11-2007 (YD)

CR w/ QC OPS CHECKS

Service Technician: [Signature]

Date: 12-04-2008