

**Customer Information:**

Company Name: Boro of Edgewood B: \_\_\_\_\_ S: \_\_\_\_\_  
 Date Received: 2-25-10 Date Given to Service: \_\_\_\_\_  
 Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

**Product Information:**

Product: 7410-7110-6510-6810-7510 Serial #: AR WJ-0017  
 9510- Drug Tester 5000  
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR -  
 Whole Instrument Top 1/2 Sim Ser #: \_\_\_\_\_  
 Other \_\_\_\_\_ Probe Ser # DD P -

**ACCESSORIES:**

110 V A/C Adapter  Regulator  Mag Card Rdr# \_\_\_\_\_  
 Printer Paper  Printer Ribbon  Casio # \_\_\_\_\_  
 Mouthpieces  Carrying Case  Dry Gas  
 Other (Please Specify) \_\_\_\_\_ Warranty Exp. Date \_\_\_\_\_

**Repair Information:**

Test #

Reason for Return: FL

Part Number	Description	Qty	Total Cost
<u>6808455</u>	<u>FL</u>	<u>1</u>	<u>n/c w/warr</u>
<u>MP CAL 21</u>	<u>cal</u>	<u>1</u>	<u>n/c w/warr</u>
<u>MP Labor</u>	<u>Labor</u>	<u>.5</u>	<u>n/c w/warr</u>

Repair Notes:

NEW FL# ARAN-0984, Cleaned spots off  
CUvette  
cal w/ QC & ops check

Service Technician: JS

Date: 03/17/2010