

# Return & Repair Form:

10381929  
80488415

### Customer Information:

Company Name: Boro of N. Plainfield B: \_\_\_\_\_ S: \_\_\_\_\_

Date Received: 12-3-09 Date Given to Service: \_\_\_\_\_

Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

### Product Information:

Product: 7410-7110-6510-6810-7510 Serial #: AR WJ-0019

9510- Drug Tester 5000  
Description: A - B - Plus - Screener - Demo Printer Ser #: AR -

Whole Instrument Top 1/2 Sim Ser #: \_\_\_\_\_

Other \_\_\_\_\_ Probe Ser # DD P -

### ACCESSORIES:

110 V A/C Adapter  Regulator  Mag Card Rdr# \_\_\_\_\_

Printer Paper  Printer Ribbon  Casio # \_\_\_\_\_

Mouthpieces  Carrying Case  Dry Gas

Other (Please Specify) \_\_\_\_\_ Warranty Exp. Date \_\_\_\_\_

### Repair Information:

Test #

Reason for Return: NEW PCB

Part Number	Description	Qty	Total Cost
<u>8315075</u>	<u>PCB</u>	<u>1</u>	<u>N/C WARR</u>
<u>MP CAL 71</u>	<u>CAL</u>	<u>1</u>	<u>N/C WARR</u>
<u>6808486</u>	<u>Plate</u>	<u>1</u>	<u>N/C WARR</u>
<u>6808455</u>	<u>FC</u>	<u>1</u>	<u>N/C WARR</u>
<u>MP Labor</u>	<u>Labor</u>	<u>.5</u>	<u>N/C WARR</u>

### Repair Notes:

Install New PCB, NEW FC ARAM-0248  
Received with NO lid.

cal w/Qc / Ops check

Service Technician: JS

Date: 12/31/09