

RETURN AND REPAIR FORM



Customer Information

B:

S:

Company Name: North Plainfield Police Dept.

Date Received: 07/17/12

Date given to service: 07/17/12

Carrier: FedEx UPS USPS

Shipping Method: GRD 3DAY 2DAY
 NDA-PRI NDA-STD

Product: 6510 6810
 7110 8610
 7510 DT5000
 7410 Upper-half 9510
 7410 Whole

Serial Number: AR WJ-0019

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: _____

Description: A B Plus Demo Screener Trade In

Accessories

110V A/C Adapter Regulator Printer Ribbon Printer Paper
 Mouthpieces 9510 Stylus 9510 Top Cover Carrying Case
 Dry Gas Other (please specify) _____

Repair Information:

Test#:

| Part Number | Description | Qty | Total Cost |
|-------------|-------------|-----|------------|
| MP Cal 71 | Calibration | 1 | NC/W |
| 6808455 | Fuel Cell | 1 | NC/W |
| MP Labor | Labor | 1 | NC/W |

Repair Notes: Unit arrived with a error 31 , Replaced defective fuel cell with fuel cell number ARDD-1065.

CAL W/QC AND OPS CHECK

Service Technician JS

Date: 07-24-12