

RETURN AND REPAIR FORM



Customer Information

B: _____

S: _____

Company Name: North Plainfield P.D., North Plainfield, NJ

Date Received: 10/11/2016

Date given to service: 10/11/2016

Carrier: FedEx UPS USPS

Shipping Method: GRD 3DAY 2DAY
 NDA-PRI NDA-STD

Product: 6510 6810
 7110 8610
 7510 DT5000
 7410 Upper-half 9510
 7410 Whole

Serial Number: AR WJ-0019

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: December 2016

Description: A B Plus Demo Screener Trade In

Accessories

110V A/C Adapter
 Mouthpieces
 Dry Gas

Regulator
 9510 Stylus
 Other (please specify) _____

Printer Ribbon
 9510 Top Cover

Printer Paper
 Carrying Case

Repair Information:

Test#: _____

Part Number	Description	Qty	Total Cost
4414161	Calibration	1	NC/W
4414166	Labor	.5	NC/W
6808455	Fuel Cell	1	NC/W

Repair Notes: Replaced defective fuel cell.

Fuel Cell = ARJC-1793

CAL W/QC AND OPS CHECK

Service Technician BC

Date: 11/02/2016