

# Return & Repair Form:

## Customer Information:

Company Name: Manasquan PD B: \_\_\_\_\_ S: \_\_\_\_\_

Date Received: 3-12-09 Date Given to Service: \_\_\_\_\_

Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

## Product Information:

Product: 7410 7110 - 6510 - 6810 Serial #: AR WJ - 0020

Description: A - B - Plus - Screener - Demo Printer Ser #: AR -

Whole Instrument Top 1/2 Sim Ser #: \_\_\_\_\_

Other \_\_\_\_\_ Probe Ser # DD P -

## ACCESSORIES:

110 V A/C Adapter  Regulator  Mag Card Rdr# \_\_\_\_\_

Printer Paper  Printer Ribbon  Casio # \_\_\_\_\_

Mouthpieces  Carrying Case  Dry Gas

Other (Please Specify) \_\_\_\_\_ Warranty Exp. Date \_\_\_\_\_

## Repair Information:

Test # \_\_\_\_\_

Reason for Return: FC

Part Number	Description	Qty	Total Cost
<u>6808455</u>	<u>FC</u>	<u>1</u>	<u>N/C W/MPR</u>
<u>MPLAC 71</u>	<u>CAL</u>	<u>1</u>	<u>N/C W/MPR</u>
<u>MPLABOR</u>		<u>.5</u>	<u>N/C W/MPR</u>

Repair Notes:

REPLACE FC  
CAL w/ QC SOPS CHECKS

Service Technician: [Signature]

Date: 03/13/2009