

RETURN AND REPAIR FORM



Customer Information

Company Name: Manasquan Boro: P.D., Manasquan, NJ

Date Received: 02/10/2017 Date given to service: 02/10/2017

Carrier: FedEx UPS USPS Shipping Method: GRD 3DAY 2DAY
 NDA-PRI NDA-STD

Product: 6510 6810 7110 8610 7510 DT5000 7410 Upper-half 9510 7410 Whole

Serial Number: AR WJ-0020
 Printer Serial#: AR
 Sim Serial#: DD
 Probe Serial#: DD

Warranty Expires: _____

Description: A B Plus Demo Screener Trade In

Accessories

110V A/C Adapter Regulator Printer Ribbon Printer Paper
 Mouthpieces 9510 Stylus 9510 Top Cover Carrying Case
 Dry Gas Other (please specify) _____

Repair Information:

Test#: _____

Part Number	Description	Qty	Total Cost
4414161	Calibration	1	NC
4414166	Labor	.5	NC

Repair Notes: Reconnected a loose wire which was causing Error 43.

CAL W/QC AND OPS CHECK

Service Technician: BC

Date: 02/15/2017