

Dräger safety

Return & Repair Form

Customer Information:

CUSTOMER NY State Police Cust # _____
 Date of Return: 6/12 Carrier: Fed Ex Method: 3 Day

Product Information:

Product: 7110 7410 6510 6810 Serial # 7110-0082
 Description: A B PLUS SCREENER _____ Printer Serial # _____
 Whole Inst. Top Half Simulator Serial # _____
 OTHER: _____ Probe Serial # _____
 Demo Unit Returned to stock on: _____
 Accessories: (check all that apply):
 110VAC Adapter Regulator Mag Card Rdr # _____
 Printer Paper Printer Ribbon Casio # _____
 Mouthpieces Carrying Case Dry Gas
 Other (specify): _____
 Warranty Expiration Date: _____

Repair Information:

Test # _____

Reason for Return: Failing SOLID-CHANGE for high results

Part #	Description	Qty	Total
<u>MPCAL71</u>	<u>7110 CAL</u>	<u>1</u>	

Repair Notes: Noted high readings on both IR + EC by about 6-7%. Calibrated unit w/ ops checks and QC. Retested unit to verify stability the next day

Technician: [Signature]

Date: 6/15/06