RETURN AND REPAIR FORM

Dräger

Customer Information

B: S:

Company Name: Fair Lawn P.D., Fair Lawn, NJ

Date Received: 05/22/13 **Date given to service:** 05/22/13

Carrier: X FedEx UPS USPS **Shipping Method:** GRD X 3DAY 2DAY
NDA-PRI NDA-STD

Product:

	6510	6810
X	7110	8610
	7510	DT5000
	7410 Upper-half	9510
	7410 Whole	

Serial Number: AR WM-0076

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: January 2014

Description: A B Plus Demo Screener Trade In

Accessories

110V A/C Adapter	Regulator	Printer Ribbon	Printer Paper
Mouthpieces	9510 Stylus	9510 Top Cover	Carrying Case
Dry Gas	Other (please specify)		

Repair Information

Test#:

Part Number	Description	Qty	Total Cost
MP Cal 71	Calibration	1	NC/W
6808455	Fuel Cell	1	NC/W
6808486	Spacer plates	2	NC/W
MP Labor	Labor	.5	NC/W

Repair Notes: Replaced defective fuel cell and spacer places.
Fuel Cell = AREB-1713

CAL W/QC AND OPS CHECK

Service Technician BC Date: 05/28/13

sustsfs010\Public\Forms\return and repair form 07_21_2011.pdf