

# Return & Repair Form:

## Customer Information:

Company Name: Richterfield B: \_\_\_\_\_ S: \_\_\_\_\_  
 Date Received: 3-8-11 Date Given to Service: \_\_\_\_\_  
 Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

## Product Information:

Product: 7410-~~4110~~6510-6810-7510 Serial #: AR WM-0079  
 9510- Drug Tester 5000  
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR -  
 Whole Instrument Top ½ Sim Ser #: \_\_\_\_\_  
 Other \_\_\_\_\_ Probe Ser # DD P -  
**ACCESSORIES:**  
 110 V A/C Adapter  Regulator  Mag Card Rdr# \_\_\_\_\_  
 Printer Paper  Printer Ribbon  Casio # \_\_\_\_\_  
 Mouthpieces  Carrying Case  Dry Gas  
 Other (Please Specify) \_\_\_\_\_ Warranty Exp. Date \_\_\_\_\_

## Repair Information:

Test #

Reason for Return:			
Part Number	Description	Qty	Total Cost
<u>10808374</u>	<u>motor</u>	<u>1</u>	<u>N/CW</u>
<u>mp cal 71</u>	<u>Cal</u>	<u>1</u>	<u>N/CW</u>
<u>mp labor</u>	<u>Labor</u>	<u>.5</u>	<u>N/CW</u>

### Repair Notes:

Unit was tested and the motor made a squeaky noise replaced motor re-test unit and no other problem found.  
Cal w/ DC ; ops check

Service Technician: [Signature]

Date: 3/10/2011