

Return & Repair Form:

Customer Information:

Company Name: Hazlet B: _____ S: _____

Date Received: 3-8-11 Date Given to Service: _____

Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

Product Information:

Product: 7410-7110-6510-6810-7510 Serial #: ARWM - 0087

9510- Drug Tester 5000

Description: A - B - Plus - Screener - Demo Printer Ser #: AR -

Whole Instrument Top 1/2 Sim Ser #: _____

Other _____ Probe Ser # DD P -

ACCESSORIES:

110 V A/C Adapter Regulator Mag Card Rdr# _____

Printer Paper Printer Ribbon Casio # _____

Mouthpieces Carrying Case Dry Gas

Other (Please Specify) _____ Warranty Exp. Date _____

Repair Information:

Test #

Reason for Return:

Part Number	Description	Qty	Total Cost
<u>808455</u>	<u>FC</u>	<u>1</u>	<u>N/A</u>
<u>mp cal 71</u>	<u>cal</u>	<u>1</u>	<u>N/A</u>
<u>mp labor</u>	<u>labor</u>	<u>15</u>	<u>N/A</u>

Repair Notes:

FC # ARWJ-0034 leaking, replaced
with FC # ARCA-0239

cal w/ qc & ops check

Service Technician: [Signature]

Date: 3/10/2011