

# Return & Repair Form:

## Customer Information:

Company Name: Verona PD B: \_\_\_\_\_ S: \_\_\_\_\_  
 Date Received: 04/10/2009 Date Given to Service: 04/10/2009  
 Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY Other

## Product Information:

Product: 7410 (7110) - 6510 - 6810 Serial #: AR WM - 0092  
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR -  
 Whole Instrument Top ½ Sim Ser #: \_\_\_\_\_  
 Other \_\_\_\_\_ Probe Ser # DD P -  
**ACCESSORIES:**  
 110 V A/C Adapter  Regulator  Mag Card Rdr# \_\_\_\_\_  
 Printer Paper  Printer Ribbon  Casio # \_\_\_\_\_  
 Mouthpieces  Carrying Case  Dry Gas  
 Other (Please Specify) \_\_\_\_\_ Warranty Exp. Date \_\_\_\_\_

## Repair Information:

Test # \_\_\_\_\_

Reason for Return: FC / IR DETECTOR

Part Number	Description	Qty	Total Cost
<u>6809458</u>	<u>IR SENSOR DETECTOR</u>	<u>1</u>	<u>N/C WARR</u>
<u>6808455</u>	<u>FUEL CELL</u>	<u>1</u>	<u>N/C WARR</u>
<u>MPCAL 71</u>	<u>CAL</u>	<u>1</u>	<u>N/C WARR</u>
<u>MP Labor</u>	<u>labor</u>	<u>.5</u>	<u>N/C WARR</u>

Repair Notes:

REPLACE IR DETECTOR  
REPLACE FC

CAL W/ QC & OPS CHECKS

Service Technician: [Signature]

Date: 04/10/2009