

Return & Repair Form:

Customer Information:

Company Name: NATIONAL PARK SERVICE B: _____ S: _____
STANDY HOOK
 Date Received: 12-29-08 Date Given to Service: 12-29-08
 Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

Product Information:

Product: 7410 - 7110 - 6510 - 6810 Serial #: ARXA - 0032
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR -
 Whole Instrument Top 1/2 Sim Ser #: _____
 Other _____ Probe Ser # DD P -
ACCESSORIES:
 110 V A/C Adapter Regulator Mag Card Rdr# _____
 Printer Paper Printer Ribbon Casio # _____
 Mouthpieces Carrying Case Dry Gas
 Other (Please Specify) _____ Warranty Exp. Date _____

Repair Information:

Test #

Reason for Return: Fuel Cell

Part Number	Description	Qty	Total Cost
<u>6B08455</u>	<u>FC</u>	<u>1</u>	<u>w/c warr</u>
<u>MP CAL 71</u>	<u>CAL</u>	<u>1</u>	<u>NIC WARR</u>
<u>MPLABOR</u>			<u>SHR N/WARR</u>

Repair Notes:

REPLACE FC / CAL w/ QC & OPS CHECKS

Service Technician: [Signature]

Date: 12-30-2008