

RETURN AND REPAIR FORM



Customer Information

B:

S:

Company Name: National Park Service, Highlands, NJ

Date Received: 03/11/14 Date given to service: 03/11/14

Carrier: FedEx UPS USPS Shipping Method: GRD 3DAY 2DAY
 NDA-PRI NDA-STD

Product: 6510 6810
 7110 8610
 7510 DT5000
 7410 Upper-half 9510
 7410 Whole

Serial Number: AR XA-0032
 Printer Serial#: AR
 Sim Serial#: DD
 Probe Serial#: DD

Warranty Expires: _____

Description: A B Plus Demo Screener Trade In

Accessories

110V A/C Adapter Regulator Printer Ribbon Printer Paper
 Mouthpieces 9510 Stylus 9510 Top Cover Carrying Case
 Dry Gas Other (please specify) _____

Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
MP Cal 71	Calibration	1	NC/W
MP Labor	Labor	8	NC/W
8315075	Alcotest 7110 MKIII Motherboard	1	NC/W

Repair Notes: Replaced defective Alcotest 7110 MKIII Motherboard.

CAL W/QC AND OPS CHECK

Service Technician BC Date: 03/14/2014